

Bid Document

Bid Details	
Bid End Date / Time	12-07-2018 19:00:00
Bid Validity	15 (Days)
Ministry Name	DELHI
Organization Name	Municipal Corporation of Delhi (MCD)
Department Name	Urban Development Department Delhi
Total Quantity	10
Item Category	blood pressure recording units
Average Turn Over of Last 3 Years	100 Lakh (s)
Experience with Gov. Required	2 Year (s)

Blood Pressure Recording Units (10)

Technical Specifications

Specification	Specification Name	Values	Allowed Values
CONFIGURATION DETAILS	Type Of SPHYGMOMANOMETERS or Blood Pressure recording unit:	Digital	Mercurial,Aneroid,Digital
	Conformity to Indian Standrad for SPHYGMOMANOMETERS	NA for digital	IS:3390 latest for Mercurial,IS 7652 latest for Aneroid,NA for digital
	Measuring device	Automated	Mechnical,Automated
	Type	Portable	Portable,Mountable/Fixed,Inclined,Integrated
	ISI MARKED	NA for digital	Yes,NA for digital
	CM/L No and validity date	N/A	-
	Measurement Method	others	-
	Display for Digital BP instrument:	NA for others	-
	Range: Pressure measurement (mmHg)	0 to 300	-
	Range:Pulse measurement for	40 to 200	-

	digital with accuracy \pm 5%. (per minute)		
	Accuracy:Pressure measurement (mmHg)	+/- 3	-
	Pressure Detection	Capacitive Pressure sensor	-
	Inflation	Automatic Pressure Application by pump	-
	Deflation and rapid Air Release	Automatic Pressure Release Valve	-
	Cuff Range	Small (18-22)cm	Small (18-22)cm,Medium (22-32)cm, Large (32-45)cm
	Dimensions of the Equipment (mmXmm)	300x300	-
	Dimensions of the Display screen (mmXmm)	300x300	-
	Operating Temperature range	0 to 55 degree Celsius	-
	Storage Temperature range	(-10) to 55 degree Celsius	-
	Automatic power off if system is idle for 3 minutes.	Yes	-
	Weight (in gram)	500	-
Power	Power Source:	AA size chargeable battery set	AA size chargeable battery set, Either Chargeable Battery Set Or {220 Volt, 50 Hz AC Supply }
	Type of AA Size chargeable Battery (AA Size)	Alkaline	Alkaline,Lithium,Ni-cd
WARRANTY, CERTIFICATIONS, TEST REPORTS	Warranty Period (in Year)	1	1,2,3
	Availability of Type Test Report covering complete test parameters as per IS3390 latest for Mercurial and IS7652 latest for Aneroid type, from any ILAC/NABL accredited/Central Gov LABFor digital type including	Yes	-

Environmental test from any ILAC/NABL accredited/Central Gov LAB		
Type test report No and date	FD(2475) 05/01/2018	-
Name and address of the Lab where test Conducted.	Amritsar Lab	-
CE Marking Certificate for B.P. Apparatus Model.	Yes	-
Conformity to OILM-R-16-1 and OILM-R-16-2 for both Non-invasive mechanical sphygmomanometers and Non-invasive automated sphygmomanometers are for guidance and All the Gazette of india related to legal metrology issued time to time, are applicable to the specification as well as to the product.	Yes	-
Others Certificate ie ROHS,WHO,GMP etc.	N/A	-
For Digital Sphygmomanometer ,Availability of Type test reports consisting of verification of all the features & functional parameters & environmental tests sequences as under: a) Dry heat test (For 16 hrs at a temp of 55degree C in accordance with IS: 9000/part-3/section-3/1977)latest	Yes	-
For Digital Sphygmomanometer Cold test(For 4 hrs at a temp of (-10) degree C in accordance with IS: 9000/part-2/section-3/1977latest c) Damp Heat Cyclic Test (For 2 cycles of 24 hrs at a temp 55 degree C &	Yes	-

	95% RH in accordance with IS: 9000/part-5/section-1/1981)latest Note:- The BP Instrument shall be checked for all the parameters before conditioning After completion of the above environmental tests sequence, with a recovery period of 1 to 2 hrs, BP Instrument shall be functional "	
--	--	--

* Bold specifications are the golden parameters.

Consignees and Quantity

S.No.	Consignee	Address	Quantity	Delivery Days
1	DR. SANJAY KUMAR	110007,Hindu Rao Hospital Malka Ganj	10	15

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Buyer Specific Additional Terms and Conditions

- 1. Scope of supply (Bid price to include all cost components) :** Only supply of Goods
- 2. Procurement under this bid is reserved for purchase from :** Micro and Small Enterprises whose credentials are validated online through Udyog Aadhaar for that product category.
- 3. Purchase preference will be given to : (if L-1 is not from selected category and Seller (s) from selected category has/have quoted price within L-1+ ----(to be specified) % of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and award contract for percentage of total value as defined/ decided in relevant policy) Micro and Small Enterprises whose credentials are validated online through Udyog Aadhaar for that product category.**
- 4. Inspection charges shall be borne by the supplier and it shall not be reimbursed by indenter to them at any point in time. Cost of inspection shall be factored in by the supplier in the cost of the product itself while submitting bid. The inspection**

agency will conduct the inspection and submit the certificate to the supplier as well as the buyers and consignees. The bidder will be solely responsible for pre-inspected material for Quality Assurance including pre-despatch inspection/testing at manufacturer's site by

5. Only BIS license holding firms are eligible to bid: Yes

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---