



Bid Number/बोली क्रमांक (बिड संख्या):  
GEM/2024/B/5507337  
Dated/दिनांक : 15-10-2024

### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	05-11-2024 09:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	05-11-2024 09:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	90 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Maharashtra
Department Name/विभाग का नाम	Public Health And Family Welfare Department Maharashtra
Organisation Name/संगठन का नाम	N/a
Office Name/कार्यालय का नाम	Office Of The Chief Executive Officer Mh Esi Soc
Total Quantity/कुल मात्रा	5656
Item Category/मद केटेगरी	medical stools or chairs (Q3) , Mobile Folding Privacy Screen / Ward Screen (Q3) , clinical cabinets (Q3) , Patient shifting rollers (Q3) , Patient Bed (Fowler and Semi Fowler) (Q3) , foot step for hospitals (Q3) , Stretcher Trolleys (Q3)
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	300 Lakh (s)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	1200 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No

**Bid Details/बिड विवरण**

<b>Document required from seller/विक्रेता से मांगे गए दस्तावेज़</b>	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Additional Doc 4 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
<b>Do you want to show documents uploaded by bidders to all bidders participated in bid?/</b>	No
<b>Past Performance/विगत प्रदर्शन</b>	70 %
<b>Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया</b>	No
<b>Type of Bid/बिड का प्रकार</b>	Two Packet Bid
<b>Primary product category</b>	medical stools or chairs
<b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b>	2 Days
<b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
<b>Evaluation Method/मूल्यांकन पद्धति</b>	Group wise evaluation
<b>Arbitration Clause</b>	No
<b>Mediation Clause</b>	No

**EMD Detail/ईएमडी विवरण**

Advisory Bank/एडवाइजरी बैंक	State Bank of India
schedule 1 EMD Amount/ईएमडी राशि (In INR)	350000
schedule2 EMD Amount/ईएमडी राशि (In INR)	150000

**ePBG Detail/ईपीबीजी विवरण**

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) /ईपीबीजी प्रतिशत (%)	3.00
Duration of ePBG required (Months)/ईपीबीजी की अपेक्षित अवधि (महीने).	15

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). The EMD Amount will be applicable for each schedule/group selected during Bid creation.

(c). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

**Beneficiary/लाभार्थी :**

Director

Office of the Chief Executive Officer MH ESI Soc, Public Health and Family Welfare Department Maharashtra, N/A,  
(Dr Shashi Govindrao Kolnoorkar)

**MSE Purchase Preference/एमएसई खरीद वरीयता**

MSE Purchase Preference/एमएसई खरीद वरीयता	No
---	----

**MII Purchase Preference/एमआईआई खरीद वरीयता**

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
5. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 70% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

**Evaluation Method ( Group Wise Evaluation Method )**

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

<b>Evaluation Schedules</b>	<b>Item/Category</b>	<b>Consignee/Reporting Officer</b>	<b>Consignee Address</b>	<b>Quantity</b>
schedule 1	Medical Stools Or Chairs	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	615
	Mobile Folding Privacy Screen / Ward Screen	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	392
	Clinical Cabinets	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	1176
	Patient Shifting Rollers	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	153
	Clinical Cabinets	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	152
	Clinical Cabinets	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	109
	Patient Bed (fowler And Semi Fowler)	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	457
	Foot Step For Hospitals	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	338
	Clinical Cabinets	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	111
	Medical Stools Or Chairs	Dr Shashi Govindrao Kolnoorkar	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	147

schedule e2	Stretcher Trolleys	Dr Shashi Govindrao Kolnoorka r	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	18 2
	Stretcher Trolleys	Dr Shashi Govindrao Kolnoorka r	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	11 1
	Stretcher Trolleys	Dr Shashi Govindrao Kolnoorka r	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	11 3
	Stretcher Trolleys	Dr Shashi Govindrao Kolnoorka r	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	21 6
	Stretcher Trolleys	Dr Shashi Govindrao Kolnoorka r	Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel 400013	84

### Medical Stools Or Chairs ( 615 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	<b>Adjustable Height</b>	Yes
	<b>Normal Height (mm)</b>	660
	<b>Min. Height (mm)</b>	660
	<b>Max. height (mm)</b>	775
	<b>Stool Top base Thickness (mm)</b>	12
Material and shape of Stool	<b>Cushioning in the stool Top</b>	with Cushion
	<b>No of Legs in the under structure</b>	5
	<b>Foot rest</b>	Without
	<b>Material of Top</b>	Wood
	<b>Painting</b>	Powder coating
	<b>Type of Stool</b>	Revolving
	<b>Back rest</b>	With

**Additional Specification Parameters - Medical Stools Or Chairs ( 615 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	<ul style="list-style-type: none"> <li>• 17" W x 20" D x 19" Seat Height &amp; 34" total height</li> <li>• The apostrophe back &amp; seat is formed in one piece with good quality rexine &amp; stitching.</li> <li>• Back &amp; Seat mounted on base base frame made by of 25mm x 1.2mm thick CRCA tubular framew fitted with PVC shoes.</li> <li>• Base Frame is powder coated in black clour</li> </ul>

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	615	15

**Mobile Folding Privacy Screen / Ward Screen ( 392 pieces )****Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERIC	<b>Partition provided</b>	With curtain
	<b>Provision of folding arrangement</b>	2 Folds, 4 Folds
	<b>Construction of middle section</b>	In 2 pcs and joined together at middle
	<b>Number of hinges(to join panels)</b>	6 Nos
MATERIAL	<b>Frame Material</b>	MS tubes of 19.05 mm (outer dia) X 1.22 mm thickness
	<b>Base material</b>	MS tubes of 25 mm (outer dia) X 1.22 mm thickness
	<b>Castors wheel</b>	Plastic castors wheel
	<b>Hinges material</b>	MS

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Hooks material	SS 304
DIMENSION	Overall height	1680 mm
	Overall width	2360 mm
	Sidepanels width (if specified)	N.A
	Middlepanels width (if specified)	N.A
CURTAIN DETAILS	Curtain type	N.A
	Curtain material	Cotton fabric
WARRANTY	Warranty period	1 - Year
CERTIFICATION	Test reports to be furnished to buyer's on demand	No
	OEM of offered product is ISO 9000: 1certified	No
	OEM of offered product is ISO 14000: 1certified	No
	OEM of offered product is ISO 18000: 1certified	No
	Product having BIFMA certification	Yes, No

**Additional Specification Parameters - Mobile Folding Privacy Screen / Ward Screen ( 392 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall approx. size: L Open 2360 mm / closed 1200 mm X W 550 mm X H 1730 mm Construction: MS CRCA tubular structure of 19 mm dia. mounted on six wheels of 50 mm dia. Two side folding.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषितो/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	392	15

### Clinical Cabinets ( 1176 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Overall depth range (mm)	800 mm (+/- 2.5%), 400 mm (+/- 2.5%)
	Overall Width range (mm)	400 mm (+/- 2.5%)
	Overall Height range (mm)	400 mm(+/-2.5%), 820 mm (+/- 2.5%)
CONFIGURATION DETAILS	Painting	Powder Coating
	Type of slide of drawer	None
	Material of Enclosure of bedside locker	Stainless steel
	Material of cabinet and Drawer	Stainless steel
	Locking system for cabinet and drawer	No
FEATURES	Number of Shutters	0
	Number of Open shelves	1
	Number of Shelves	2
	Finish	Power Coated
FRAME TUBE	Frame Tube Material	STAINLESS STEEL

#### Additional Specification Parameters - Clinical Cabinets ( 1176 pieces )



Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	: 400mm W x 400mm D x 800mm H (Approx.) Construction: MOC MS CRCA & 304 Stainless steel top. Main cabinet should be made of 1 mm thick "C" shaped one-piece box. Cabinet should have one drawer at top & cabinet at bottom and open space in between drawer & box. Door should be made up of 1 mm thick CRCA sheet pivoted at top & bottom by 4 mm diameter pin. Door & drawer should have inbuilt handle. The drawer should have sliding channel.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013, Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	1176	15

**Patient Shifting Rollers ( 153 pieces )**

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम कटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
General features	Frame material	Stainless Steel, Aluminium
	Foldable Type	No Foldable

**Additional Specification Parameters - Patient Shifting Rollers ( 153 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
------------------------------	----------------------------------

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall Dimension: 2035 mm L x 620 mm W x 880 mm H (Approx.) Stretcher top approx. size: 2030 mm L x 520 mm W" Construction: Frame should be made up of 50 mm x 25 mm x 18G CRCA rectangular tube, upper horizontal support tube should be made up of 25 mm x 25 mm x 18G tube. Four no's 150 mm diameter X 40 mm W heavy duty metal body Swiveling casters bolted to an outer pipe of 31.75 x 1.2 mm diameter.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	153	15

**Clinical Cabinets ( 152 pieces )**

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	<b>Overall depth range (mm)</b>	900 mm (+/- 2.5%), 400 mm (+/- 2.5%)
	<b>Overall Width range (mm)</b>	400 mm (+/- 2.5%), 850 mm(+/-2.5%)
	<b>Overall Height range (mm)</b>	820 mm (+/- 2.5%)
CONFIGURATION DETAILS	<b>Painting</b>	Powder Coating
	<b>Type of slide of drawer</b>	None
	<b>Material of Enclosure of bedside locker</b>	Stainless steel, Mild steel
	<b>Material of cabinet and Drawer</b>	Mild steel
	<b>Locking system for cabinet and drawer</b>	Yes
FEATURES	<b>Number of Shutters</b>	1
	<b>Number of Open shelves</b>	4
	<b>Number of Shelves</b>	4

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Finish	Power Coated
FRAME TUBE	Frame Tube Material	MILD STEEL

#### Additional Specification Parameters - Clinical Cabinets ( 152 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	"Overall Dimension : 36" L x 18" W x 72" H MOC Mild Steel CRCA Sheet & Tubes Main frame 25mm x 25mm x 1.2mm tube CRCA tubular sheet with side paneling made up of 22 gage CRCA sheet ""4 no Adjustable shelves should be made up of 20 gage CRCA sheet. Door made up of tubular frame work fitted with Glass 1.5mm thick It should have locking facility

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	152	15

#### Clinical Cabinets ( 109 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	Overall depth range (mm)	900 mm (+/- 2.5%), 400 mm (+/- 2.5%)
	Overall Width range (mm)	400 mm (+/- 2.5%), 850 mm(+/-2.5%)
	Overall Height range (mm)	820 mm (+/- 2.5%)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
CONFIGURATION DETAILS	Painting	Powder Coating
	Type of slide of drawer	Fitted With Telescopic Channel
	Material of Enclosure of bedside locker	Wood
	Material of cabinet and Drawer	Mild steel
	Locking system for cabinet and drawer	Yes
FEATURES	Number of Shutters	1
	Number of Open shelves	1
	Number of Shelves	1
	Finish	Power Coated
FRAME TUBE	Frame Tube Material	MILD STEEL

#### Additional Specification Parameters - Clinical Cabinets ( 109 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall Size: 610 mm L x 355mm D x 1535 mm H Construction :- Front Frame should be made up of 50 mm x 25 mm x 1.2mm thick CRCA sheet & back side frame should be 25 mm x 25 mm x 1.2 mm CRCA sheet Front frame & back frame are joined together by 50mm x 25mm x 1.2mm thick CRCA pipe The Back & front 1mm thick CRCA sheet

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	109	15

## Patient Bed (Fowler And Semi Fowler) ( 457 pieces )

### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Performance Parameters	<b>Type of bed</b>	Semifowler
	<b>Number of bed platform sections</b>	2 for semi fowler bed, 4 for fowler bed
	<b>Availability of Leg section</b>	NA for semifowler
	<b>Leg section linkage</b>	NA for semifowler
	<b>Manual adjustments</b>	Backrest through two screw systems with thrust bearings individually manouvred by handle for semifowler bed
	<b>The bottom end of the H legs are provided with PVC shoes</b>	Yes in case of bed without castors
	<b>Material of handle</b>	SS
	<b>Provision of castors</b>	NA (No castors provided)
	<b>Number of hooks in Telescopic I V Pole</b>	2
	<b>Number of sections in mattress</b>	NA (incase mattress is not provided )
	<b>Provision of collapsible Railing</b>	No
Material and Dimensional Parameters	<b>Size of Buffer in mm</b>	100
	<b>Overall length of Back Rest</b>	825-850
	<b>Overall length of Knee Rest</b>	NA for semifowler
	<b>Overall width of Knee Rest</b>	NA for semifowler
	<b>Overall width of Leg Rest</b>	NA for semifowler
	<b>Overall length of Fix Section</b>	100-150
	<b>Overall width of fix section</b>	825-850
	<b>Colour of mattress cover</b>	NA (incase mattress is not provided )
	<b>Material of collapsing railing</b>	NA (incase railing not provided)
	<b>Types of collapsing railing</b>	NA (incase railing not provided)
Reports	<b>Availability of test report from govt/ NABL/ILAC accredited lab covering all parameters</b>	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes

#### Additional Specification Parameters - Patient Bed (Fowler And Semi Fowler) ( 457 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Construction: The Bed frame should be made up of 60 mm X 30 mm X 16G CRCA rectangular tube with four sleeves of diameter 40 mm X 3 mm thick, 100 mm long welded at four corners. Four sections top should be made up of 18G CRCA sheet; double press bend on four sides and uniformly embossed holes from 25mm to 15mm dia. and embossing depth of 4mm in four rows and distance between holes should be 125mm apart.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	457	15

#### Foot Step For Hospitals ( 338 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Design	Type of Foot Step	Double Foot Step
	Height of Foot Step	150 mm for first step and 355 mm for second step from the ground for Double Foot Step

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimension	Length of Foot Step (mm)	451-500
	Width of Foot Step (mm)	451-475
	Thickness of Sheet for Top (mm)	0.8-0.9
	Size of Pipe (mm)	20
	Thickness of Pipe (mm)	1.2-1.3
Material	Material of Sheet of Top	Mild Steel
	Mild Steel Pipe for Frame Structure	Square
Finish	Finish	Powder coated
	Colour of Paint	Ivory White

#### Additional Specification Parameters - Foot Step For Hospitals ( 338 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	280mm/ 510 mm step height and size approx 460 mm L x 300 mmW MS tubular construction with 18 g CRCA sheet. Steps super imposed by aluminum chucked plate & legs fitted with rubber feet. Finish: All components should be pretreated in separate eight tank process for better finish, good adhesion and corrosion protection.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013, Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	338	15

#### Clinical Cabinets ( 111 pieces )

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
DIMENSIONS	<b>Overall depth range (mm)</b>	340 mm (+/- 2.5%), 900 mm (+/- 2.5%)
	<b>Overall Width range (mm)</b>	375 mm (+/-2.5%), 850 mm(+/-2.5%)
	<b>Overall Height range (mm)</b>	750 mm (+/- 2.5%)
CONFIGURATION DETAILS	<b>Painting</b>	Powder Coating
	<b>Type of slide of drawer</b>	Fitted with Ball bearing type
	<b>Material of Enclosure of bedside locker</b>	Stainless steel
	<b>Material of cabinet and Drawer</b>	Mild steel
	<b>Locking system for cabinet and drawer</b>	No
FEATURES	<b>Number of Shutters</b>	0
	<b>Number of Open shelves</b>	1
	<b>Number of Shelves</b>	1
	<b>Finish</b>	Power Coated
FRAME TUBE	<b>Frame Tube Material</b>	STAINLESS STEEL

**Additional Specification Parameters - Clinical Cabinets ( 111 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall Dimension: 38" L x 18"H x 78"H MOC MS CRCA Sheet & Tubes. Should have a welded construction for a sturdier frame. The cupboard should have 15 compartments with doors having external lock fitting arrangement. Construction: Structure should be made of 25.4 mm x 25.4 x 18 G MS ERW tube & 22 guage CRCA Sheet. Door should be made of 20 guage MS CRCA Sheet with slot for fixing lock. It should be provided with PVC leveller.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**



S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	111	15

### Medical Stools Or Chairs ( 1447 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	<b>Adjustable Height</b>	Yes
	<b>Normal Height (mm)</b>	600
	<b>Min. Height (mm)</b>	600
	<b>Max. height (mm)</b>	700
	<b>Stool Top base Thickness (mm)</b>	75
Material and shape of Stool	<b>Cushioning in the stool Top</b>	with Cushion
	<b>No of Legs in the under structure</b>	4
	<b>Foot rest</b>	With
	<b>Material of Top</b>	Wood
	<b>Painting</b>	Polishing
	<b>Type of Stool</b>	Revolving
	<b>Back rest</b>	Without

#### Additional Specification Parameters - Medical Stools Or Chairs ( 1447 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall approx. size: 380mm x 380mm x H 515mm. Top approx. size: 300mm x 300mm. Construction: MOC MS CRCA pipe and SS 304 grade sheet. The top should be made up of 304 grade stainless steel sheet 0.9mm thick double press bent on four sides. The edges of the stainless steel sheet should be folded to 180 degree to avoid the sharp edges. The stainless steel top should be fixed with the help of pop revits in slush position.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	1447	15

**Stretcher Trolleys ( 182 pieces )**

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	Overall Length of trolley in mm with $\pm 5\%$ tolerance	2030
	Overall width of trolley in mm with $\pm 5\%$ tolerance	710, 560
	Material of Stretcher Trolley	Stainless Steel
	Material of railing	Stainless steel, Not provided
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1

**Additional Specification Parameters - Stretcher Trolleys ( 182 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Specification: Overall approx. dimension: 515mm L x 595mm W x 1180mm H. (including handle) Construction: Main frame should be made up of MS CRCA rectangular combination of 50mm x 25mm x 18 g and 25mm x 25mm x 18g. The Trolley should have 4 no's tray for keeping medicine. The Top tray should be made up of 304 grade stainless steel sheet of 20 g and remaining 3 no's tray should be made up of MS CRCA Sheet of 20 gauge.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Additional Specification Documents/अतिरिक्त विशिष्टि दस्तावेज़**

Applicable Specification Document	<a href="#">View</a>
-----------------------------------	----------------------

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	182	15

**Stretcher Trolleys ( 111 pieces )****Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	Overall Length of trolley in mm with $\pm 5\%$ tolerance	2030
	Overall width of trolley in mm with $\pm 5\%$ tolerance	560
	Material of Stretcher Trolley	Mild Steel
	Material of handle	SS

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1

#### Additional Specification Parameters - Stretcher Trolleys ( 111 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Specification: Overall approx. dimension: 1040 mm L x 500 mm W x 850 mm H. Construction: Frame should be made up of 25 x1.2 mm thick stainless steel tube. Top and bottom frame should be made up of 19 mm x 18G stainless steel tube

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Additional Specification Documents/अतिरिक्त विशिष्टि दस्तावेज़

Applicable Specification Document	<a href="#">View</a>
-----------------------------------	----------------------

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	111	15

#### Stretcher Trolleys ( 113 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	Overall Length of trolley in mm with $\pm 5\%$ tolerance	1905, 2030
	Overall width of trolley in mm with $\pm 5\%$ tolerance	710
	Material of Stretcher Trolley	Stainless Steel
	Type of railing	Tuck down type railings, Not provided
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	3

#### Additional Specification Parameters - Stretcher Trolleys ( 113 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall approx. dimension: 710 mm L x 460 mm W x 850 mm H. Construction: MOC 304 grade stainless steel. Vertical and horizontal pipes should be welded to form a sturdy frame. Vertical Frame should be made up of pipe 25.24 x 1.2 mm round thick stainless-steel tube. horizontal pipes 19.25 mm x 1mm 18G round stainless-steel tube. Both shelves should be made up of 24G thick 304

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Additional Specification Documents/अतिरिक्त विशिष्टि दस्तावेज़

Applicable Specification Document	<a href="#">View</a>
-----------------------------------	----------------------

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	113	15

### Stretcher Trolleys ( 216 pieces )

#### Technical Specifications/तकनीकी विशिष्टियाँ

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	<b>Overall Length of trolley in mm with <math>\pm 5\%</math> tolerance</b>	1905
	<b>Overall width of trolley in mm with <math>\pm 5\%</math> tolerance</b>	710
	<b>Material of Stretcher Trolley</b>	Stainless Steel
	Material of railing	Stainless steel
WARRANTY	<b>Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b>	1, 5

#### Additional Specification Parameters - Stretcher Trolleys ( 216 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	"Overall approx. dimension Height : 790mm. Outer Base diameter 440 mm. ID145mm Pull Type Oxygen Trolley should be made from 1" dia. MS CRCA tubular framework mounted on 50 mm dia Castors."

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Additional Specification Documents/अतिरिक्त विशिष्टि दस्तावेज़

Applicable Specification Document

[View](#)

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	216	15

**Stretcher Trolleys ( 84 pieces )**

**Technical Specifications/तकनीकी विशिष्टियाँ**

[\\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
PRODUCT INFORMATION	Overall Length of trolley in mm with $\pm 5\%$ tolerance	2030
	Overall width of trolley in mm with $\pm 5\%$ tolerance	710
	Material of Stretcher Trolley	Stainless Steel
	Material of railing	Stainless steel, Not provided
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	1

**Additional Specification Parameters - Stretcher Trolleys ( 84 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
Buyer added Specific Requirement	Overall Size: 925mm L x 650mm D x 1480mm H Five Shelves approx. dimension: 610mm L x 510mm D Drawer approx.: dimension: 580mm L x 440mm D Construction: Structure should be made up of MS CRCA, 25mm x 50mm vertical channels & 50mm x 25mm MS ERW 1.2mm square tube, mounted on 125mm dia. swiveling castors two with brake""

\* Bidders offering must also comply with the additional specification parameters mentioned above.

#### Additional Specification Documents/अतिरिक्त विशिष्टि दस्तावेज़

Applicable Specification Document	<a href="#">View</a>
-----------------------------------	----------------------

#### Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dr Shashi Govindrao Kolnoorkar	400013,Office of the Chief Executive Officer, Maharashtra Employees State Insurance Society Panchdeep Bhavan, 6th Floor, N.M. Joshi Marg, Opp. Western Railway Workshop Lower Parel	84	15

#### Special terms and conditions-Version:1 effective from 18-10-2022 for category medical stools or chairs

1.

- 1. Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
- 2. Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
- 3. Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
- 4. Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or



otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.

5. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
6. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
7. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
8. **Software:** All software updates should be provided free of cost during warranty period.
9. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

#### **Special terms and conditions-Version:2 effective from 18-10-2022 for category clinical cabinets**

##### **1. 1. Comprehensive warranty:**

Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

##### **2. Service centres:**

Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

##### **3. Source of supply:**

It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them .

##### **4. Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods

should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.

#### **5. Spare Parts:**

Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied

#### **6. Installation, Training, Manuals:**

Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.

#### **7. Electrical safety checking:**

Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

#### **8. Software:**

All software updates should be provided free of cost during warranty period.

### **Special terms and conditions-Version:2 effective from 12-01-2024 for category Patient shifting rollers**

1. 1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc if applicable. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc if applicable.
3. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
4. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.

## Special terms and conditions-Version:4 effective from 26-06-2024 for category Patient Bed (Fowler and Semi Fowler)

1.
  1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
  2. The sellers are registered on GeM based on self-declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
  3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
  4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
  5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
  6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
  8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
  10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.  
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
  11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation &

commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.

12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.
14. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

### **Special terms and conditions-Version:1 effective from 18-10-2022 for category foot step for hospitals**

1.

1. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
2. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
3. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
4. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
5. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
6. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and

tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.

7. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
8. **Software:** All software updates should be provided free of cost during warranty period.
9. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

### **Special terms and conditions-Version:1 effective from 21-03-2024 for category Stretcher Trolleys**

1. 1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Medical Device License (if applicable), product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.

### **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

#### **1. Generic**

**OPTION CLAUSE:** The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

#### **2. Generic**

**Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

#### **3. Generic**

**Consortium:** In case of Contracts, wherein the seller alone does not have necessary expertise, the seller

can form consortium with other sellers for submission of the bid, with one of the consortium company as leader. However, each and every member of the consortium shall be equally responsible for the complete execution of the project contract. An undertaking to this effect is to be uploaded with bid.

4. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

5. **Generic**

**End User Certificate:** Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.

6. **Generic**

**Shelf Life:** The Product/Spare parts to be supplied as part of the services must have minimum

36 Months of  
Shelf Life. On the date of supply, minimum  
36 Months  
usable shelf life should be available / balance.

7. **Generic**

The successful bidder has to supply all essential accessories required for the successful installation and commissioning of the goods supplied. Besides standard accessories as per normal industry practice, following accessories must be part of supply and cost should be included in bid price:

Transport , Inspection , Installation

8. **Generic**

1. The Seller shall not assign the Contract in whole or part without obtaining the prior written consent of buyer.
2. The Seller shall not sub-contract the Contract in whole or part to any entity without obtaining the prior written consent of buyer.
3. The Seller shall, notwithstanding the consent and assignment/sub-contract, remain jointly and severally liable and responsible to buyer together with the assignee/ sub-contractor, for and in respect of the due performance of the Contract and the Sellers obligations there under.

9. **Generic**

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

- i) The Seller fails to comply with any material term of the Contract.
- ii) The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
- iii) The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
- iv) The Seller becomes bankrupt or goes into liquidation.
- v) The Seller makes a general assignment for the benefit of creditors.
- vi) A receiver is appointed for any substantial property owned by the Seller.
- vii) The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

10. **Turnover**

**Bidder Turn Over Criteria:** The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a

certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

**11. Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

**12. Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

**13. Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

**14. Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

**15. Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

**16. Certificates**

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

**17. Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

MAHARASHTRA EMPLOYEE STATE INS SOC  
payable at  
MUMBAI

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

**18. Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

**19. Turnover**

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

## 20. Certificates

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

## Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।



---Thank You/धन्यवाद---