



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/5555969
Dated/दिनांक : 28-10-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	08-11-2024 19:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	08-11-2024 19:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Odisha
Department Name/विभाग का नाम	Home Department Odisha
Organisation Name/संगठन का नाम	Odisha Police
Office Name/कार्यालय का नाम	State Police Headquarters
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	Raman Spectrophotometer System (Forensic use) (Q2)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	1400 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	1 Year (s)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Additional Doc 4 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	Yes
Past Performance/विगत प्रदर्शन	10 %

Bid Details/बिड विवरण	
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	Yes
RA Qualification Rule	H1-Highest Priced Bid Elimination
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	5 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	700000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

AIG of Police (Provisioning)
State Police Headquarters, Home Department Odisha, Odisha Police , Cuttack
(Bamadev Singh)

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	No
---	----

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

5. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

Pre Bid Detail(s)

Pre-Bid Date and Time	Pre-Bid Venue
29-10-2024 18:00:00	The Prospective bidders are requested to submit their queries relating to the tender/bid of the AI through Mail ID – dirsfsi.orpol@nic.in by 04.11.2024 by 05.00 PM. The queries will not be consider after stipulated period. In this connection a pr e-bid technical committee will be held in virtual mode and the link will be sent through mail for the meeting. The pre-bid date is tentatively nature and subjected to postponement which be duly in formed through mail

Raman Spectrophotometer System (Forensic Use) (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
CONFORMITY/TYPE	Conformity to technical requirements/Governing specification	As per Approved Specification of Directorate of LNJJN for ' RAMAN SPECTORPHOTOMETER SYSTEM WITH INTEGRATED COLOR VIDEO MICROSCOPE WITH OTHER SUITABLE ACCESSORIES, vide letter No. 14/37/2019 -LNJJN NICFS Dated 7th October, 2019(Copy of Approved Specification uploaded in GeM portal)
	Data Library	Narcotics & drugs Library, Explosives Library, Preloaded Ink Library with facility to built new database of ink etc as per governing specification for "Raman Spectrophotometer
WARRANTY	Warranty	12, 24 Or higher (month)
ACCESSORIES	ACCESSORIES	Suitable accessories to attach it to microscope, NA

Additional Specification Parameters - Raman Spectrophotometer System (Forensic Use) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Data Library	i. Hydrocarbons ii. Psychoactive Substances & Illicit Drugs iii. Dyes, Pigments & Stains iv. Facility to built new database of ink etc as per governing specification for "Raman Spectrophotometer"
ACCESSORIES	Suitable accessories to attach it to microscope
Test report of complete product to prove the conformity of product as per specification	OEM test report

* Bidders offering must also comply with the additional specification parameters mentioned above.

Additional Specification Documents/अतिरिक्त विशिष्टि दस्तावेज़

Applicable Specification Document	View
-----------------------------------	----------------------

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shyamamani Sahu	751010,Rasulgarh, Bhubaneswar	1	30

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Raman Spectrophotometer System (Forensic use)

Sl. No.	Parameters	Value
1.	Gem Category	Raman Spectrophotometer System (Forensic use) (Q2 Category)
2.	Conformity to technical requirements: Governing specification	As per Approved Specification of Directorate of LNJN for 'RAMAN SPECTROPHOTOMETER SYSTEM WITH INTEGRATED COLOR VIDEO MICROSCOPE WITH OTHER SUITABLE ACCESSORIES, vide letter No. 14/37/2019 -LNJN NICFS Dated 7th October, 2019(Copy of Approved Specification uploaded in GeM portal)
3.	Data Library	<ul style="list-style-type: none"> i. Hydrocarbons ii. Psychoactive Substances & Illicit Drugs iii. Dyes, Pigments & Stains iv. Facility to built new database of ink etc as per governing specification for "Raman Spectrophotometer"
4.	ACCESSORIES	Suitable accessories to attach it to microscope
5.	Test report of complete product to prove the conformity of product as per specification	OEM test report

Additional Parameters / corrigendum for Raman Spectrophotometer System

Sl. No.	Specifications	Value
---------	----------------	-------

1.	Raman System	The Raman system should be developed specially for forensic examinations and must include dedicated casework management software providing automatic spectral comparisons and data archiving with search and match facilities. It should provide a comprehensive library of forensic Raman spectra for identification of unknowns.
2.	Laser Probe	<p>i. It must have external cavity stabilized laser system running at 785 ± 5 nm; with Laser power up to 80m W at source or better including all safety; It must be a class 3B laser product or better.</p> <p>ii. 1 nos. of EXTRA compatible Laser probe must be supplied with the system by the OEM/Supplier/Bidder.</p>
3.	Spectral Resolution	It must have Spectral resolution of minimum 6-7 wave numbers or better.
4.	Spectral Range	It must have Spectrometer range of 200 - 3200 wave numbers or better range.
5.	Laser Power Controls	It should provide adjustable operation of maximum power.
6.	Detector	It must have Thermoelectrically Cooled CCD array detector.
7.	Video Microscope	It should have Microscope turret capable of holding microscope objectives: 5x, 10x and 20x or better objective fitted as standard; Colour video camera to align laser spot on sample; A4 size X-Y-Z manipulation stage with ± 5 mm of X-Y movement at sample centre.
8.	Software	Must include Software package for spectral analysis, including fluorescence removal filters & comparison casework reports: Chemometric routines, such as PCA, PQS, SNV, MSC; Comprehensive library data basing facility; Instrument calibration routine and calibration standard; Software updates for 10 years. Compatible dedicated software to operate integrated Raman Spectrophotometer.

9.	Raman Database(Full-time Perpetual Licensed Version for all)	<p>Must include Forensic Raman Spectra Collections database</p> <p>a) 86 Raman spectra of compounds commonly of interest in the forensic laboratory.</p> <p>S.T. Japan Forensic Raman Spectra Library</p> <p>b) 748 Raman spectra of compounds commonly of interest in the forensic laboratory.</p> <p>S.T. Japan Complete Collection Database</p> <p>Comprehensive Collection contains 31 S.T. Japan Raman Application Spectra Databases</p>
10.	Library for Raman Spectrometer(Full-time Perpetual Licensed Version for all)	<p>Forensic Raman spectra library, Forensic Database, New Psychoactive Substances and illicit Drugs / Narcotics Drug, Insecticide and pesticide, Hydrocarbons, Pharmaceuticals Drugs, Antibiotics, Hazardous and Toxic Chemicals , Polymer and polymer Additives, Alcohols and Phenols, Esters, Lactones and Anhydrides, Hydrocarbons, Pesticides, Semiconductor Chemicals, Flavours, Fragrances and Cosmetics, Dyes, Pigments, Stains, Food Additives and Food Packaging, Solvents, Biochemicals, Aldehydes and Ketones, , High Production Volume Chemicals, Minerals, Sulphur and Phosphorus compounds, , Inorganics materials, Excipients, White Powders, Coatings, Paints, Selected Organic Compounds, EuroPharm, Environmental Pollutants, Essential oils, Petrochemicals, Gemstones, Organometallics & silicon compounds, Microplastics & Related compounds and Common Contaminants. (Perpetual Licensed Version for all)</p>
11.	UPS for Raman Spectrophotometer system	<p>UPS for Raman Spectrophotometer system</p> <p>i. 3 KVA or more Online UPS suitable for the “Raman Spectrophotometer system” must be of a reputed manufacturer.</p> <p>ii. UPS from reputed manufacturer for uninterrupted (real-time) power supply with a backup time of minimum 30 minutes for PC. UPS must come with in-built isolation transformer maintenance free batteries.</p> <p>UPS must come with three years or more warranty and 5 years or more warranty on batteries from the date of installation.</p>

12.	PCs, UPS, Printer and software	<p>Computer system with minimum 5years onsite warranty and must be following minimum configuration. A standalone PC should be there for system control with suitable PC, Monitor size 24" or more and compatible colour laser Printer for printing of reports / results along with 3 sets of additional printing cartridges.</p> <p>PC must be instrument control and data acquisition software.</p>
13.	Warranty	<p>5 years from the date of installation for all parts. It includes CMC. If any break down on laser within warranty period it must be replaced once without any charge.</p>
14.	General Accessories	<p>Complete set of accessories for a fully working system</p>
15.	INSTALLATION AND TRAINING	<ul style="list-style-type: none"> i. Application training must be done at the place of install by the OEM onsite / online mode with the presence of technical team. ii. The Raman Spectrometer system should be supplied in prescribed time period, installed by trained technical person, demonstrated and operational training to be imparted to Scientists and technical officers (minimum 5 or more persons) at State F.S.L. Bhubaneswar Chemistry laboratory till satisfactory. iii. Calibration certificates, Warranty certificates and installation certificate, as applicable, shall be provided along with the supply of the item. <p>One sets of hard copies of the manuals descriptions along with soft copy shall be supplied along with the Raman Spectrometer system.</p>
16.	GENERAL TERM & CONDITION	<ul style="list-style-type: none"> i. The above equipment has to be supplied along with its all accessories, installed, and commissioned for functional and performance testing at SFSL BHUBANESWAR Laboratory. ii. The offered system should be complete and all essential cables, interfaces with Life span certificate of the equipment and other parts should be supplied with the system. iii. System support non-destructive methodology for forensic document examination. iv. Equipment should be supplied with suitable calibration material as per ASTM. v. Offer must clearly specify list of standard accessories supplied with the main Instrument. vi. Maintenance to be performed by a OEM service

		<p>ce engineer, or trained technical person.</p> <p>vii. Supplier must support annual maintenance of the instrument after expiry of the warranty of the instrument.</p> <p>viii. Must provide PBG (performance bank guarantee) for 3 years.</p> <p>ix. In case of instrument breakdown OEM / supplier / bidder must supply all the consumables and accessories of the instrument free of cost for 5 years if required without any conditions.</p> <p>x. OEM / Supplier / Bidder must check and rectify the power supply like earthing and supply current before installation of the entire equipment.</p> <p>xi. All the Quoted technical specification should be available on manufactures website/catalogue to judge the authenticity of technical specification. Separate sheet will not accept.</p> <p>xii. 10 years assurance for providing spares and other requirement for running the supplied model of Raman Spectrometer system in State F.S.L. Bhubaneswar.</p> <p>xiii. Comprehensive warranty (on-site) and maintenance (with factory trained engineer) for minimum 5 year for all parts including labour with co-maintenance and service and free shipping on all warranty items and also for regular up gradation to all software during the entire warranty period from the date of installation.</p> <p>xiv. Supplier / Bidder should have produce the past performance certificate of the similar item from reputed laboratory / Forensic laboratories / Govt. organization.</p> <p>xv. Response time to attend any fault of this warranty shall 48 - 72 hrs.</p>
--	--	---

1. Bidder Annual Turnover for last three years should not be less than 700 Lakhs.

2. Bidder are requested to submit ITR of Last 3 Years.

3. The Prospective bidders are requested to submit their queries relating to the tender/bid of the AI through Mail ID - dirtsfl.orpol@nic.in by 04.11.2024 by 05.00 PM. The queries will not be consider after stipulated period. In this connection a pre-bid technical committee will be held in virtual mode and the link will be sent through mail for the meeting. The pre-bid date is tentatively nature and subjected to postponement which be duly informed through mail

4. This bid is governed by OGFR-2023 and GFR regulations. Bidders will liable for punitive measures, including potential debarment, outlined in OGFR and GFR

for any violations or failure to comply with GFR / OGFR as well as bid terms and conditions.

5. The bidders are requested to submit EMD / e-BG as per bid requirement in term of electronics Bank Guarantee quoting Unique Identification Number (UIN) " NCO GD2383P" with NeSL while applying eBG from their respective Banks.

Manufacturer's Authorization Form (MAF)
(To be submitted on OEM Letterhead)

Ref. No. _____ Dated _____

To

The A.I.G. of Police (Provisioning),
Odisha, Cuttack. Dear Sir,

Tender Reference No./ NIT /GeM Bid No. : _____

Name of the item:

We M/s. _____ who are established and reputable manufacturers of _____ do here by authorize M/s. _____ 6 / 9 _ (Name and address of Agent / Dealer) to participate in the above tender.

We hereby extend our technical assistance to the bidder during installation and inspection of the product.

We hereby certify that, the Equipments being sold would not be declared End of Support (EoS) or become obsolete in the next 5 years. Also, we certify that the products being sold would be covered under Warranty / Support and OEM support will be available for ___ years(as specified in the GeM Bid / Tender / NIT No.) from the date of installation, even in the case, the bidder becomes "Out of service".

We have studied the requirements of the product and confirm that we will adhere to the specifications of the tender and quality plan and extend all support during the inspection and provide documentary evidence at the time of inspection for the verification by Client/Client's representative.

Yours faithfully
(Name)
For and on behalf of M/s. _____
(Name of manufacture)

AIG of Police (Provisioning) Odisha, Cuttack.

2. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

3. **Inspection**

Nominated Inspection Agency: On behalf of the Buyer organization, any one of the following Inspection Agency would be conducting inspection of stores before acceptance:
Pre-dispatch Inspection at Seller Premises (applicable only if pre-dispatch inspection clause has been selected in ATC):

No

Post Receipt Inspection at consignee site before acceptance of stores:
Yes

4. **Forms of EMD and PBG**

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C (Name of the Buyer). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

5. **Warranty**

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

6. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

7. **Certificates**

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

8. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

9. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

10. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

AIG of Police (Provisioning) odisha Cuttack

A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

11. Certificates

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---