



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/5589697
Dated/दिनांक : 11-11-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	23-11-2024 20:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	23-11-2024 20:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Odisha
Department Name/विभाग का नाम	Home Department Odisha
Organisation Name/संगठन का नाम	Odisha Police
Office Name/कार्यालय का नाम	State Police Headquarters
Total Quantity/कुल मात्रा	3
Item Category/मद केटगरी	Hydrogen Generator (Q3) , Nitrogen Generator (Q3) , Zero Air generator with compressor (Q3)
OEM Average Turnover (Last 3 Years)/मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)	280 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	1 Year (s)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Additional Doc 4 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No

Bid Details/बिड विवरण	
Past Performance/विगत प्रदर्शन	10 %
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	Yes
RA Qualification Rule	H1-Highest Priced Bid Elimination
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	Hydrogen Generator
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	5 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	140000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	62

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

AIG of Police (Provisioning)
State Police Headquarters, Home Department Odisha, Odisha Police , Cuttack
(Bamadev Singh)

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	No
---	----

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	No
--	----

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

5. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

Hydrogen Generator (1 pieces)**Technical Specifications/तकनीकी विशिष्टियाँ**

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
---------------	------------------------------------	--

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Technical Specifications	Flow rate of hydrogen gas (cc/min)	100-500
	Purity of Hydrogen Gas	99.9999, 99.99996 (percent)
	Material	Stainless Steel, Mild Steel
	delivery Pressure	5-100 psig +/-0.5 psig
Power Details	Voltage	230.0 (Volt)
	Frequency	50-60 Hz
Features	Additional Features	Water management system, Remote Control, proton exchange membrane, Automatic loading pump, Portable, Noiseless Performance, Others, Silent operation with minimal maintenance
	Safety Features	Low water level audible alarm, Automatic Shutdown, Others, Leak detection with auto shutdown, safety check up system
Environment Conditions	Temperature	40.0 (degree Celsius)

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shyamamani Sahu	751010,Rasulgarh, Bhubaneswar	1	30

Nitrogen Generator (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Technical Specification	Membrane Life Time-Min	15 Or higher (year)
	Nitrogen Storage Tank	Any applicable numeric value (liter)
	Number of Air Compressors	1, 2, 3, 4 Or higher

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Capacity in Liter/minute	500.0 Or higher
	Features	Earthing Piping, Zero Air Generator for low investment, Low Electric power consumption, Calibrated needle valve to adjust purge flow, Low Maintenance, Online Oxygen display (ppm or %)and Service Alarm
Safety Parameter	Safety Valve	Yes
Additional Feature	Mist Eliminator	Yes
	Safety Pressure Relief Valve	Yes
	Alarm Shutdown	Yes
	Automatic Drain Trap	Yes
Warranty Specification	Warranty	Any applicable numeric value (year)

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Shyamamani Sahu	751010,Rasulgarh, Bhubaneswar	1	30

Zero Air Generator With Compressor (1 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Technical Parameter	Startup Time	5.0 (minute)
	Purity	GC/UHP Grade

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Shyamamani Sahu	751010,Rasulgarh, Bhubaneswar	1	30

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Additional Specification

Additional Parameters / corrigendum for Hydrogen Gas Generator	
Purity of Hydrogen Gas (%)	Minimum purity should be 99.9999
Material	Mild Steel / Steel
Packaging Type	Wooden Box
Flow rate	Minimum 600 ml/min
Delivery Pressure	Upto 10 bar adjustable

Tech nology	Proton Exchange membrane cell (PEM)
Cell life	Should be more than 8 years
Operations	Separate controller with touch screen which allows monitoring of operational status, all report alarms & functioning parameters remotely.
Environmental moisture	≤ 80%
Water Tank	should have easily removable and filling water without needing to stop generator
Electronic Display	<p>All report, alarms & Process parameters should be displayed on remote Control. Which can be interfaced with the CPU (Remote Control) or PCs for remote control.</p> <p>ii. Hydrogen pressure will be displayed on screen of CPU or equivalent technology to monitor.</p> <p>iii. Generator is having log book facility with software and service due indicator in built in the system.</p> <p>iv. To avoid over pressure within generator, generator gets auto shutoff in case of excess pressure or equivalent technology to auto shutoff.</p> <p>v. Service due indicator inbuilt in system</p>
Additional Parameters / corrigendum for Nitrogen Gas Generator	
Features	Earthing Piping, Nitrogen Generator for low investment, Low Electric power consumption, Calibrated needle valve to adjust purge flow, Low Maintenance and Service Alarm.
Capacity in Liter::Minute	Minimum 500

Nitrogen Storage Tank (L)	Minimum 05 liter
Purity of Nitrogen Gas (%)	Minimum purity should be 99.9995
Material	Mild Steel / Steel
Flow rate	Minimum 500 ml/min
Delivery Pressure	Minimum 5 bar or more
Technology	Pressure Swing Adsorption Technology (PSA)
Operations	Two Stage Filtration. Programmable Logic Control Panel Screen with monitor display.
Environmental moisture	≤ 80%

Electronic Display	ii.	Separate controller with touch screen which allows monitoring of operational status, all report alarms & functioning parameters remotely. All report, alarms & Process parameters should be displayed on remote Control.
	iii.	Nitrogen generator should have catalyst chamber / inbuilt Hydrocarbon removal facility.
	iv.	Outlet Nitrogen should have Hydrocarbon Content < 0.1 ppm for highly sensitive application.
	vi.	To avoid over pressure within generator, generator gets auto shutoff in case of excess pressure or equivalent technology to auto shutoff.
	v.	Service due indicator inbuilt in system

Additional Parameters / corrigendum for Zero Air Generator

Total Hydrocarbon residual	< 0.1 ppm
Hydrocarbon removal facility	Methane >99.0% ; Aromatics (BTX) & Hydrocarbons >99.5 % ; Carbon Monoxide (CO), Hydrogen (H ₂), Ethylene & Propylene > 99.9 %
Technology	Catalytic Platinum/Palladium reactor for HCTs removing Or Catalyst chamber to remove hydrocarbons (as methane) to <0.05ppm.
Output Pressure	upto 10 bar
Flow rate	Minimum 5000 ml/min or more
Delivery Pressure	Minimum 5 bar / 80 Psi or more

Elect ronic Displ ay	Controller with touch screen which allows monitoring of operational status & functioning parameters. All process parameters should be displayed.
ii.	Zero air generator should have catalyst chamber / inbuilt Hydrocarbon removal facility.
iii.	To avoid over pressure within generator, generator get auto shutoff in case of excess pressure.
iv.	Service due indicator inbuilt in system

Air Compressor

Capa city	Minimum 20 Ltr/min
Press ure	upto 10 bar
Tech nolo gy	It should be Oil free piston
Oper ation s	Separate controller with touch screen which allows monitoring of operational status, all report alarms & functioning parameters remotely.
Elect ronic Displ ay	All report, alarms & Process parameters should be displayed on remote Control.
Safet y	With all safety alarms
Inde x of Prote ction	IP 20 or better
Oper ating temp eratu re	5 °C - 45 °C
Envir onm ental mois ture	≤ 80%

Noise level	< 55 dB
Service due indicator	Inbuilt in system
Power	Should not use more power than 85VA
<u>Additional Term and Conditions (ATC)</u>	
Compatibility	All generators i.e. Hydrogen (H ₂), Nitrogen (N ₂), Zero air and Air compressor must be compatible with each other and also compatible with existing Gas Chromatography Mass Spectrometer (make: Agilent Technologies) and Gas Chromatography (Make: Thermo fisher Scientific) of Chemistry Division, State F.S.L. Bhubaneswar.
Warranty (Including CM C)	5 years from the date of installation of all generators i.e. Hydrogen (H ₂), Nitrogen (N ₂), Zero air and Air compressor. It includes CMC. Warranty should cover with two preventive maintenance visits per year and breakdown as per requirements.
Room Renovations (If required)	All the civil work / electrical earthing and electrifications are in the scope of the OEM/ Bidder /Supplier. All the accessories must be supplied by the OEM / Bidder / Supplier at the time of installation. The offered system should be complete and all essential cables, interfaces and other parts should be supplied with the system.
Installation and training	Installation and demonstration to be done by supplier or OEM. IQ (Installation Qualification) / OQ (Operational Qualification) and PQ (Performance Qualifications) should be done at the place of install by the supplier. iii. The generator system should be supplied in prescribed time period, installed by trained technical person, demonstrated and operational training to be imparted to Scientists and technical officers (minimum 5 persons) at State F.S.L. Bhubaneswar Chemistry Division till satisfactory.

General terms & conditions	<p>WPM (Working Procedure Manual) must be provided by the vendors.</p> <p>All the gas generator must be supplied from the same make so that it can be for compatible to each other..</p> <p>All the generator are preferably stackable to save the space.</p> <p>iv. Hard copies of the manuals with circuit diagrams and descriptions along with soft copy shall be supplied along with the supplied system.</p> <p>v. One set of tools required for maintenance of the items must be supplied.</p> <p>vi. All the Quoted technical specification should be available on manufactures website/catalogue to judge the authenticity of technical specification.</p> <p>vii. To maintain quality standard and safety Gas generator must Certified by CE and Index of Protection IP 20 or better.</p> <p>viii. Calibration certificates, Warranty certificates and installation certificate, Equipment Life Span Certificate as applicable, must be provided along with the supply of the item.</p> <p>ix. Response time to attend any fault of the above items shall 48 – 72 hrs.</p> <p>x. Service- should have all the spares stock in Local service office. Also assurance of minimum 10 years for providing spares and other requirement for running the supplied model of Generators in State F.S.L. Bhubaneswar.</p> <p>xi. Air compressor must be compatible with supplied hydrogen, Nitrogen and zero air generator to SFSL Bhubaneswar and also compatible with present GCMS (make: Agilent Technologies) and GC (Make: Thermo fisher Scientific) of Chemistry Division, State F.S.L. Bhubaneswar.</p> <p>xii. OEM / Supplier / Bidder must supply all the consumables free of cost including installation kit, possible parallel connection and four point pressure regulator kit etc. to run the generator for 5 years.</p> <p>xiii. One time shifting must be performed by the OEM with IQ,OQ and PQ free of cost , if required any.</p> <p>xiv. Minimum 05 nos. of installation in any Forensic Laboratories and 05 nos. of installation in other Government Organization.</p>
----------------------------	---

–

ADDITIONAL Term Conditions

1. The bidder must submit ITR last three years.
2. Bidder Average annual turnover should be 254 lakhs.
3. This bid is governed by OGFR-2023 and GFR regulations. Bidders will liable for punitive measures, including potential debarment, outlined in OGFR and GFR for any violations or failure to comply with GFR / OGFR as well as bid terms and conditions.

4. The bidders are requested to submit EMD/e-BG as per bid requirement in term of electronics Bank Guarantee quoting Unique Identification Number (UIN) "N COGD2383P" with NeSL while applying eBG from their respective Banks.

Manufacturer's Authorization Form (MAF)

(To be submitted on OEM Letterhead)

Ref. No. _____ Dated _____

To

The A.I.G. of Police (Provisioning),
Odisha, Cuttack.

Dear Sir,

Tender Reference No./ NIT / GeM Bid No. : _____

Name of the item:

We M/s. _____ who are established and reputable manufacturers of _____
_____ do hereby authorize M/s. _____
_____ (Name and address of Agent / Dealer) to participate in the above tender.

We hereby extend our technical assistance to the bidder during installation and inspection of the product.

We hereby certify that, the equipments being sold would not be declared End of Support (EoS) or become obsolete in the next 5 years. Also, we certify that the products being sold would be covered under Warranty / Support and OEM support will be available for ____ years (as specified in the GeM Bid / Tender / NIT No.) from the date of installation, even in the case, the bidder becomes "Out of service".

We have studied the requirements of the product and confirm that we will adhere to the specifications of the tender and quality plan and extend all support during the inspection and provide documentary evidence at the time of inspection for the verification by Client/Client's representative.

Yours faithfully

(Name)

For and on behalf of M/s. _____

(Name of manufacturers)

2. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

3. Inspection

Nominated Inspection Agency: On behalf of the Buyer organization, any one of the following Inspection Agency would be conducting inspection of stores before acceptance:
Pre-dispatch Inspection at Seller Premises (applicable only if pre-dispatch inspection clause has been selected in ATC):

No

Post Receipt Inspection at consignee site before acceptance of stores:

Yes

4. Forms of EMD and PBG

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C (Name of the Buyer). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

5. **Warranty**

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

6. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

7. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

8. **Certificates**

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

9. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

10. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

11. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

AIG of Police (Provisioning) Odisha Cuttack

A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

12. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---