



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/5596584
Dated/दिनांक : 11-11-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	02-12-2024 11:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	02-12-2024 11:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	90 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Railways
Department Name/विभाग का नाम	Indian Railways
Organisation Name/संगठन का नाम	South Western Railway
Office Name/कार्यालय का नाम	South Western Railway
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	Anaesthesia Workstation (V2) (Q2)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC), OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Comprehensive Maintenance Charges Required	Yes
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	Yes

Bid Details/बिड विवरण

Inspection to be carried out by Buyers own empanelled agency	Yes
Type Of Inspection	Pre Dispatch
Name of the Empanelled Inspection Agency/ Authority	RITES
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	56520

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	38

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

PFA/SWR/Hubli
SOUTH WESTERN RAILWAY, RAIL SAUDHA, GADAG ROAD, HUBLI -580020
(Pfa/swr/hubli)

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

MSE Purchase Preference/एमएसई खरीद वरीयता

1. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

2. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

3. Inspection of Stores by Nominated Inspection Authority / Agency of buyer or their authorized representatives

An independent third party Professional Inspection Body can help buyer in mitigating buyer's risk with pre-dispatch/post-dispatch inspection in order to ensure that equipment, components, solutions and documentation conform to contractual requirements. The buyer has a right to inspect goods in reasonable manner and within reasonable time at any reasonable place as indicated in contract. Inspection Fee/ Charges (as pre-greed between buyer and Inspection Agency) would be borne by the buyer as per their internal arrangement but may be recovered from the seller if the consignment failed to conform to contractual specification and got rejected by the Inspection Officer .If so requested and accepted by the seller , initially seller may pay for inspection charges as applicable and get the same reimbursed from buyer if consignment accepted by the Inspecting Officer . For reimbursement seller has to submit proof of payment to Inspection Agency.

Seller/OEM shall send a notice in writing / e-mail to the Inspecting officer / inspection agency specifying the place of inspection as per contract and the Inspecting officer shall on receipt of such notice notify to the seller the date and time when the stores would be inspected. The seller shall, at his own expenses, afford to the Inspecting officer, all reasonable facilities as may be necessary for satisfying himself that the stores are being and or have been manufactured in accordance with the technical particulars governing the supply. The decision of the purchaser representative /inspection authority regarding acceptance / rejection of consignment shall be final and binding on the seller.

The Seller shall provide, without any extra charge, all materials, tools, labour and assistance of every kind which the Inspecting officer may demand of him for any test, and examination, other than special or independent test, which he shall require to be made on the seller's premises and the seller shall bear and pay all costs attendant thereon.

The seller shall also provide and deliver store / sample from consignment under inspection free of charge at any such place other than his premises as the Inspecting officer may specify for acceptance tests for which seller/OEM does not have the facilities or for special/ independent tests.

In the event of rejection of stores or any part thereof by the Inspecting officer basis testing outside owing to lack of test facility at sellers premises, the seller shall, on demand, pay to the buyer the costs incurred in the inspection and/or test. Cost of test shall be assessed at the rate charged by the Laboratory to private persons for similar work.

Inspector shall have the right to put all the stores or materials forming part of the same or any part thereof to such tests as he may like fit and proper as per QAP/governing specification. The seller shall not be entitled to object on any ground whatsoever to the method of testing adopted by the Inspecting officer.

Unless otherwise provided for in the contract, the quantity of the stores or materials expended in test will be borne by seller.

Inspecting officer is the Final Authority to Certify Performance / accept the consignment. The Inspecting officer's decision as regards the rejection shall be final and binding on the seller.

The seller shall if so required at his own expense shall mark or permit the Inspecting officer to mark all the approved stores with a recognised Government or purchaser's mark.

Anaesthesia Workstation (V2) (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GENERAL	Product Description	Anesthesia Workstation
	Components of Anesthesia workstation	Anesthesia machine , Inbuilt ventilator , vaporisor , Gas delivery system , Multi para monitor
Gas Delivery System	Availability of Pin index yokes for gases	Oxygen and Nitrous oxide
	Number of Yokes, Oxygen	1
	Number of Yokes, N2O	1
	NIST /DISS Inlet pipeline with Hoses for O2, N2O and Air	Yes
	Availability of flow meters for O2, N2O and air	Yes
	Type of Flowmeter	Rotameter
	Mode of gas mixing	Pneumatic
	Availability of pressure gauges for cylinder and pipelines in anesthesia workstation	Yes
	Availability of audible and visual oxygen failure alarms	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Minimum emergency oxygen flow (in liter per minute)	35 - 70
	Type of hypoxic guard with automatic cutoff of N2O	Mechanical
	Minimum 25% O2 across all Oxygen - Nitrous Oxide Mixture	Yes
	Auxillary Oxygen Flowmeter	Yes
	Auxillary Common Gas outlet for use in open circuit	Yes
Vaporizer	Number of quick mount type vaporizer for easy interchangeability at a time	2
	Number of selectatec back bar	1
	Facility of interlock for vaporizer	Yes
	Minimum agent capacity for free volatile anesthetic agent (in ml)	225 ml
	Sensor connection shall be internal to prevent disconnection	Yes
	Vaporizer compatible to gases	Sevoflurane, Isoflurane, Desflurane, Halothane
	The vaporizer design should be maintenance free	Yes
	Vaporizer calibration requirement	After every 5 years
	Should have a drug specific Filling arrangement	Yes
Breathing System	Breathing system should have fresh gas de-coupled / compensation , closed circle absorber system	Yes
	Range of pressure relief valve (in mbar) of breathing system	5 to 75
	Number of steps in change over from Mechanical / Spontaneous to bag ventilation	Single step
	Number of reusable autoclavable closed silicone circuit supplied	2 Nos for adult
	Number of Disposable closed circuits - Adult	NA
	Number of Disposable closed circuits - Pediatric	NA

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Number of Disposable closed circuits - Neonatal	NA
	Number of bair circuit (adult) supplied with workstation	5
	Number of Jackson Rees circuit for pediatric patients supplied with workstation	5
	Volume capacity of breathing system (in liters)	1.5 or more
	Reusable flow sensors covered in warranty	No
	Dual flow sensor at inspiratory and expiratory port	No
	Breathing system should have leak and compliance test	Yes
	Breathing system should be autoclavable	Yes
	Facility of CO2 Bypass	No
Anesthesia Ventilator	Available modes of operating ventilator	Manual/spontaneous, Volume controlled, Pressure controlled, SIMV/PCV, SIMV/VCV, CMV
	Ventilator should be capable to pressure support	Yes
	Ventilator should be capable to dual mode of ventilation	Yes
	Integrated ventilator not require bellows to be change for adults and infants	Yes
	Tidal volume of ventilator (in ml)	10 to 1400 ml
	PEEP of ventilator in mbar	0 to 20 mbar
	Breathing frequency in BPM	4 to 60 BPM
	Inspiratory pause percentage of Ti	0 - 50%
	Screen size of Ventilator monitor in inches	6 to 10, 10.1 to 15, 15 .1 and more Or higher
	Integrated ventilator shall be able to provide facility to monitor FiO2, dual agent, CO2, Automatic identification of agents, Minimal Alveolar Concentration (MAC) Value	Yes, No

Specification	Specification Name/विशिष्ट का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Integrated ventilator shall be able to provide facility to monitor expired O2 (EtO2)	No
	Flow Trigger Sensitivity	Adult, Pediatric, Neonatal
Monitor	Type of monitor LCD/LED/TFT	Yes
	Size of monitor in inches	12 to 15, 15.1 to 18, 18.1 and more Or higher
	Monitor should not require lengthy start up procedure or calibration for display of parameters	Yes
	The graphical and numerical trend with split screen duration	24 hours
	Patient multipara monitor shall be able to display parameters	HR, SpO2
	Capnography to be displayed	Yes
	IBP1 and IBP2 to be displayed	Yes
	CO (cardiac output) to be displayed	No
	Anesthesia Gas Monitoring (AGM)	No, Yes
	BIS/ENTROPY to be displayed	No
	Neuro Muscular Transmission(NMT) to be displayed	No
	ECG monitoring with standard lead configuration, I,II,III, avL, avF&V filters so that it is protected against defibrillator, electro surgery potentials, arrhythmia analysis and ST segment analysis with trends	Yes
	NIBP should be determine with oscillometric method, should read systolic, diastolic and mean arterial pressure and pulse rate	Yes
	Pulse oximetry with sensors suitable for adults, pediatric & neonate	Yes
	ETCO2 measurement should be by infra -red absorption technique,measuring mode should be side streamor main stream and response time should be less than 3 seconds	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Monitor should be able to measure and display all parameters in single screen	Yes
	Monitor should be provided from the same OEM	Yes
Alarms	Alarms	Tidal Volume, Minute Volume, Airway Pressure Low and High, Inspiratory Oxygen concentration, Audio power supply fail alarm, Fail to cycle warning (apnea).
	Alarms shall be adjustable to high / low limits with audio and visual alarms	Yes
Power supply	Length of main cable in meter with power socket of standard indian make to be provided	?5
	Inbuilt battery backup for monitor in Mins	120, 180, 240, 300, 360 Or higher
	Power supply	220-240V AC, 50 Hz single phase
Additional Accessories	Transparent re-usable silicon face mask size 0,1, 2, 3, 4 & 5	2 each
	Breathing bag capacity - 2 Each	1 Litre
	3/5 ECG lead	2
	Nasopharyngeal/rectal and skin Temperature probes should be Provided	2 Nos Adult
	NIBP Reusable Large adult, Reusable small adult cuffs, Reusable Pediatric cuff	2 each
	ETCO2 sample line	10
	Pressure regulator with 3 meter high Pressure tubing(1 for oxygen and 1 for Nitrous oxide)	Yes
	Provision for laryngoscopes	Not Provided
	Number of Reusable patient circuits	Not provided
	SpO2 sensor with cable provided	2 Nos Adult
	Reusable IBP interface cables	Not provided
	Disposable IBP transducers	Not provided
	BIS sensors	Not provided
	Accessories for wall mounting	NA (If trolley mounting)
ADDITIONAL PARAMETERS	Type of mounting for the Anesthesia workstation	Trolley Mounting

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Material used for Anesthesia workstation	Epoxy powder painted steel
	Mode of integrated scavenging system with pressure relief valve	Passive
	Minimum Battery backup for anesthesia workstation	1 hour
	Facility of manual ventilation in case of electricity and battery failure	Yes
	Facility of gas and agent delivery in case of electricity and battery failure	Yes
	All the components likes anesthesia machine, integrated ventilator, vaporizer from the same manufacturer	Yes
	The workstation should have one or more drawers to keep accessories, good mobility, anti-static caster wheels with locking facility & conveniently placed handles for easy movement of the machine	Yes
	All software updates should be provided free of cost during Warranty period	Yes
	The anesthesia machine can be used safely up to 1000 gauss line while being used with MRI	No
	Pressure relief	5 to 70 cm H2O
CERTIFICATIONS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Certification for manufacturing unit	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date	Yes
	Electrical Protection certification	IEC-60601-1-1 or equivalent BIS

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement	Yes
WARRANTY	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	3, 4, 5 Or higher (year)

Comprehensive Maintenance	
Warranty of required product	3 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

Installation Commissioning and Testing (ICT) details for the above item:

% of Product Cost Payable on Product Delivery	80%
Min Cost Allocation for ICT as a % of product cost	20%
Number of days allowed for ICT after site readiness communication to seller	30 Days/दिन

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Harish Shanmukhappa Tadakanahalli	580020,O/o Medical Director, Central Hospital, Gadag Road, Hubli 580 020	1	60

Special terms and conditions-Version:1 effective from 01-07-2024 for category Anaesthesia Workstation (V2)

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.

10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.
- 14.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

2. Warranty

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

3. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

- A) Bidder(s) should submit the signed copy of the following criteria along with the bid:

Sl. No	Criteria	Compliance by bidder /OEM (Yes/No)	Document attached or Not
1	The Bidder should be an OEM or Authorized agent/dealer of OEM only. In case bidder's participated as dealer/agent they Should submit bid Specific Authorization letter from OEM along with the bid.		

2	Bidder should submit the compliance of the technical Specifications including additional technical specifications mentioned in Para(C) of ATC with seal and signature of the OEM along with the bid.		
3	Original technical brochure of the product offered should be enclosed along with bid.		
4	It should have CE/US FDA certification for the product they offered or equivalent Indian standard(s). Firm should submit the valid certificate(s) along with the bid		
5	<p>Bidder should submit declaration of OEM for confirmation of supply of spares of the machine offered for a period of at least 10 years from the date of supply of the machine to SWR. Firm should submit the details</p> <p>(Name, Address, contact details, email-ID) of the service centre in India. Firm should have dedicated service engineers(s)/personnel in the city of Hubli / Dharwad / Bangalore / Mysore. Name, Address, Contact details, email-ID of the service engineer should be provided to consignee at the time of supply of machine.</p>		
6	Bidder should have successfully supplied the tendered or similar Anaesthesia Workstation with Monitor to Indian Railways or any other government organizations/PSUs against the orders in the last five completed financial year and the current financial year up to the date of tender closing. Firm must submit documentary proof such as order copy, Receipt Note/Inspection Certificate/Installation and Commissioning reports/ Completion report etc. along with the bid to substantiate their claim.		

7	Performance Certificate(s): The bidder shall, in their offer, submit performance certificate of at least 01 no of such machine supplied in the last 05 years and the machine covered by these certificate must have worked satisfactorily for at least one year from the date of commissioning as on the original date of closing of tender. The Certificate shall not be older than one year from the original date of closing of bid. The performance certificate issued after original date of closing of tender (in cases where tender closing date has been extended) are also acceptable however the machine must have completed one year of satisfactory working after date of commissioning as on original date of closing of tender. The performance certificate submitted by the bidder shall have been issued by the actual end user organization of the machine, with their clear signature and address therein, in whose premises the machine is installed and commissioned.		
8	Bidder should submit bid specific make in India Local content declaration letter for the offered product from OEM along with the bid.		

All the participating tenderers/bidders should comply all the above parameters without fail. The offer received without compliance of above parameters is liable to be ignored.

B) Bidder(s) should submit the declaration for compliance of code of integrity:

Code of Integrity Declaration : I/We hereby certify that no other concerns or affiliates (such as having common partner/director/promoter/owner/holding company /management control etc.) have participated in this tender which give rise to a conflict of interest or fall under anti-competitive practices. We understand that the offers of all such related concerns including ours are liable to be rejected as per relevant provisions of IRIS terms and conditions. I/We, affirm that we shall abide by the Code of Integrity in general, with specific reference to conflict of interest [Para 25 02 (e)], anti competitive practices and obligation to proactive disclosures. I/We understand that we are fully responsible for the above declaration and it shall not be the Railway's responsibility to verify the same. I/We are aware that we shall be liable for all consequences of violation of Code of Integrity, if detected at any stage.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent

Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

- 1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC. Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- 2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
- 3.GST shall be included in the CMC Charges quoted.
- 4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula. A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
- 5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
- 6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available. Under this option up to 10 years can be chosen for CMC charges beyond warranty period.
- 7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

- CMC charges for 1st year after warranty period- Percentage to be indicated- A1
- CMC charges for 2nd year after warranty period- Percentage to be indicated- A2
- CMC charges for 3rd year after warranty period - Percentage to be indicated- A3
- CMC charges for 4th year after warranty period - Percentage to be indicated- A4
- CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

- (i) Number of years for which CMC required.
- (ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$ and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5 shall depend on how many years CMC selected. For 3 year CMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall be matched individually.

12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाही

का आधार होगा।

---Thank You/धन्यवाद---